


HOW TO PAIR THE CARESENS DUAL BLOOD GLUCOSE METER WITH THE SMARTLOG APP ON YOUR MOBILE PHONE



1. On Phone

Download SmartLog App from either the Apple Store or Google App. Ensure that the Bluetooth setting on your phone is switched on



2. On Meter

Press and hold the **S** for 3 seconds to enter the SET Mode.

3. On Meter

Press the **←** or **→** button to select 'YES' and press the **S** button to confirm the change from 'No' to 'Yes' in bottom left of the screen.

4. On Meter

The screen will display 'bT'. Press the **→** button. The meter screen shows 'OFF', 'On', and 'PAIR' in turn. To pair your meter with your smartphone, press the S button when 'PAIR' blinks on the screen

Note:

The  symbol will appear on the meter screen when the Bluetooth feature is on.

5. On Phone

Open the SmartLog App.

Press '**More**' at right hand base of screen.



6. On Phone

Select '**Accessories**' and then select the CareSens Dual meter from the list of devices as the meter to be paired.



7. On Phone

Following the steps on the screen;

Press '**Register**'

Select '**Next**'

Select '**Search**'

8. On Phone

Select your meter on the SmartLog screen. This is identified by the last 4 digits of the meter serial number which is on the back of the meter.



9. On Meter

The meter will now be displaying a six-digit PIN number

10. On Phone

Enter this six-digit PIN number into the screen on the SmartLog mobile app and press Pair.

11. On Phone

The meter will display 'SUCCESS' once your meter and smartphone are paired and connected.



Please Note:

If, after pairing, the blood glucose readings are expressed as mg/dl in the SmartLog App on your phone, this can be changed in the '**More**' menu by selecting 'User Information' and changing the Glucose units to mmol/L

