



Visit the webpage below to see supported devices/ download app (scan the QR code for easy access).

[www.i-sens.com/smartlog](http://www.i-sens.com/smartlog)

**SPiR!T**<sup>®</sup>  
HEALTHCARE

**Spirit Healthcare Ltd.**

Spirit House, Saffron Way,  
Off Saffron Lane, Leicester, LE2 6UP  
Customer Support Tel : 0800 881 5423  
[www.spirit-healthcare.co.uk](http://www.spirit-healthcare.co.uk)

24/7 NON-AUTOMATED FREE SUPPORT  
**0800 881 5423**

**i-SENS, Inc.**  
43, Banpo-daero 28-gil, Seocho-gu,  
Seoul 06646, Korea  
[www.i-sens.com](http://www.i-sens.com)

EC REP Medical Technology Promedt  
Consulting GmbH,  
Altenhofstrasse 80, D-66386  
St. Ingbert, Germany

© 2017 i-SENS, Inc. All Rights Reserved. PGC8E2002 REVx 07/2017 CE 0123 IVD

Owner's Booklet

**TEE2<sup>+</sup>**

Blood Glucose Monitoring System



Large LCD



No Coding



Intuitive Meal Icons



**SPiR!T**<sup>®</sup>  
HEALTHCARE

Bluetooth

## **Welcome to the TEE2 Plus Blood Glucose Monitoring System**

Thank you for choosing the TEE2 Plus Blood Glucose Monitoring System. The system provides you with rapid and convenient blood glucose *in vitro* (i.e., outside the body) diagnostic monitoring. You can obtain accurate results in just 5 seconds with a small (0.5 µL) blood sample.

Test results can be sent to a smartphone through Bluetooth® wireless connection.

### **Bluetooth® wireless technology**

Bluetooth wireless technology is used by some smartphones and many other devices. TEE2 Plus Meter uses Bluetooth wireless technology to pair\* and to send your glucose results to a smartphone.

The TEE2 Plus Meter is designed to work with the SmartLog™ mobile app. When using the TEE2 Plus System, we suggest you pair your TEE2 Plus Meter with your smartphone and track your results.

\* The process of creating a connection between two Bluetooth devices. An auto-generated passkey has to be exchanged between the two devices. Once the devices are paired, they will automatically communicate with each other when the Bluetooth feature is activated.

- No part of this document may be reproduced in any form or by any means without the prior written consent of i-SENS.
- The information in this manual is correct at the time of printing. However, i-SENS reserves the right to make any necessary changes at any time without notice as our policy is one of continuous improvement.